

F7

(1971)
(Rev. 1
1989)
(Rev.2
May 1999)

Portable instruments for measuring oxygen and flammable vapour concentrations

Every oil tanker is to be provided with at least two portable gas detectors capable of measuring flammable vapour concentrations in air and at least two portable O₂ analysers.

In addition, for tankers fitted with inert gas systems, at least two portable gas detectors are to be capable of measuring concentrations of flammable vapours in inerted atmosphere.

